


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 23 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130286✓
Name		STAN SWETZOF	ID# 5002✓
Date		7/21/10✓	
A	Agency	North pole police Dept	Phone # 488-0612
Instrument Location 125 Snowman Lane, North pole, AK 99705			
Alco S/N		X124662✓	Target Value .087✓
High Pressure		900 PSI	
B	Alco Test Values	.084✓ 1 st Alco	.083✓ 2 nd Alco
Signature 			
(OVER)			

CDB
7/23/10


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 9th day of AUGUST, 2010.


Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130286

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130286

JULY 21, 2010

OPERATOR'S NAME:

SMETZOF/STANLEY

OPERATOR'S NUMBER: 5002

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: NPL1

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .007

ALCO S/N: K124662

--- BREATH ANALYSIS ---

.007 ADJUSTED FOR 29.57 in
ALCO TARGET .005 12:47
BLANK TEST .000 12:48
INTERNAL STANDARD VERIFIED 12:48
ALCO TO 29.57 in .004 12:49
BLANK TEST .000 12:49
SUBJECT SAMPLE .000 12:50
BLANK TEST .000 12:51
ALCO TO 29.57 in .003 12:51
BLANK TEST .000 12:52

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130286

JULY 21, 2010

TIME 12:43

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.57 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMN0PQRSTUWXYZ\]^_`abcdefgijklmno
pqrstuvwxyz{|}~